

**INDIAN INSTITUTE OF TECHNOLOGY ROPAR
DEPARTMENT OF CHEMISTRY**

Dated: 14-12-2009

SUBJECT: N.I.Q. for purchase of High pressure autoclave

Sealed quotations are invited for the purchase of High pressure autoclave. The equipment is to be installed at IIT Ropar (Roopnagar) Punjab. Delivery of the equipment shall be required within 1-2 months of placement of the purchase orders. The quotations should reach in the office of Dr. Rajendra Srivastava (Room No. 204, Chemistry Department), IIT Ropar, Nangal Road, Rupnagar-140001, Punjab, latest by 3:00 PM on January, 08, 2009.

Sr. No.	Name of Equipment and Brief Specifications	Quantity
1.	250 ML High pressure Lab autoclave Working pressure 80-100 Kg/cm ² Working temperature maximum limit = 250 deg. C. Zero leakage Magnetic drive coupling 4 bladed mechanical stirrer with variable speed Good quality pressure guage, vent valve, safety rupture disc, gas inlet-valve, liquid sampling valve, very sensitive thermocouple, P.I.D temperature controller with high temperature alarm system.	1

Terms and Conditions:

1. The equipments should be supplied and installed at IIT Ropar (ROOPNAGAR) Punjab.
2. The prices quoted should be inclusive of all taxes and CIF IIT Ropar.
3. The quotation received after due date will not be considered.
4. The Institute reserves the right to reject any quotation without assigning any reasons.
5. The warranty of the equipments should be for minimum 2 years from the date of installation.
6. The payment terms should be specified in the quotation clearly

(Dr. Rajendra Srivastava)