



No. 1077-17/CH-10005/Instt/PS/

16.04.2018

CORRIGENDUM

(Tender Notice No. 1070-17/CH-10005/Instt/PSdated 23.03.2018)

Reference Tender Notice No. 1077-17/CH-10005/Instt/PS/ for “High Pressure Isocratic HPLC upgradable to Binary Gradient”. The following changes have been made in the specifications:

In the RI detector and Pump

Flow rate = 0 to 10 mL/min instead of 0 to 20 mL/min

Drift = 1×10^{-6} or better instead of 1×10^{-7} or better

Software database: System should be provided with branded software such as SQL/ACCESS/ORACLE or similar database.

Warranty: Comprehensive warranty for three years including hardware consumables hardware spare part.

Installation and User list: All vendors must attach the supply order copies from state and government university/institute of last two years. Vendor must have installed 100 or more such systems in the state/government University and reputed industries.

Bidders are requested to submit their bids as per revised Technical Specifications. The last date of receipt and opening of Bids for “High Pressure Isocratic HPLC upgradable to Binary Gradient” is hereby extended upto 01.05.2018. Those who have already sent/submitted their bids can revise the same clearly indicating "Revised Bid for High Pressure Isocratic HPLC upgradable to Binary Gradient" on sealed cover.

Timings and all other details and terms & conditions remain the same.

Registrar