



CORRIGENDUM

Reference Tender Notice No. 1415-19/CV-10168/ISIRD/PS for “Advanced Soil Testing Equipment”. The following changes/additions have been made in the specifications:

Point 2. Flexible Wall Permeability Apparatus

2. b) Volume change measuring device: at least 250 cm³ with minimum resolution of 0.01 cm³

Read as

2.b) Volume change measuring device: at least 200 cm³ with minimum resolution of 0.01 cm³

Point 4. Unsaturated Triaxial Testing

1. b) Double walled cell with minimum capacity of 3 MPa, acrylic cell should have sufficient valves for hydraulic connections.

4. c) Volume change measuring device: 250 cm³ or above with min. resolution of 0.01 cm³

Read as

1. b) Double walled cell with minimum capacity of 2 MPa, acrylic cell should have sufficient valves for hydraulic connections.

4. c) Volume change measuring device: 200 cm³ or above with min. resolution of 0.01 cm³.

The last date of receipt and opening of Bids for “Advanced Soil Testing Equipment” is hereby extended upto 31.10.2019. Bidders are requested to submit their bids as per revised Technical Specifications. Those who have already sent/submitted their bids can revise the same clearly indicating "Revised Bid for “Advanced Soil Testing Equipment” on sealed cover.

Timings and all other details and terms & conditions remain the same.

Registrar