



**Government
eProcurement
System**

eProcurement System Government of India

Tender Details

Date : 25-Aug-2021 05:31 PM

Print

Basic Details

| | | | |
|---|--------------------------------------|--|--------|
| Organisation Chain | Indian Institute of Technology Ropar | | |
| Tender Reference Number | 1628-21 | | |
| Tender ID | 2021_IITRP_644129_1 | | |
| Tender Type | Global Tenders | Form of contract | Supply |
| Tender Category | Goods | No. of Covers | 2 |
| General Technical Evaluation Allowed | No | ItemWise Technical Evaluation Allowed | No |
| Payment Mode | Not Applicable | Is Multi Currency Allowed For BOQ | Yes |
| Is Multi Currency Allowed For Fee | No | Allow Two Stage Bidding | No |

Cover Details, No. Of Covers - 2

| Cover No | Cover | Document Type | Description |
|----------|-----------------------|---------------|---|
| 1 | Fee/PreQual/Technical | .pdf | Tender for Supply and Installation of Cryogenfree Transport Set up with Accessories |
| 2 | Finance | .xls | Tender for Supply and Installation of Cryogenfree Transport Set up with Accessories |

Tender Fee Details, [Total Fee in ₹ * - 0.00]

| | | | |
|-------------------------------------|------|-----------------------|-----|
| Tender Fee in ₹ | 0.00 | | |
| Fee Payable To | Nil | Fee Payable At | Nil |
| Tender Fee Exemption Allowed | No | | |

EMD Fee Details

| | | | |
|------------------------|-------|---|-----|
| EMD Amount in ₹ | 0.00 | EMD through BG/ST or EMD Exemption Allowed | No |
| EMD Fee Type | fixed | EMD Percentage | NA |
| EMD Payable To | Nil | EMD Payable At | Nil |

Foreign Currency BOQ Details

| S.No | Currency | Conversion rate in ₹ |
|------|-------------------|----------------------|
| 1 | US Dollar | NA |
| 2 | European Euro | NA |
| 3 | British Pound | NA |
| 4 | Australian Dollar | NA |
| 5 | Swiss Franc | NA |
| 6 | Japanese yen | NA |
| 7 | Canadian Dollar | NA |
| 8 | Singapore Dollar | NA |

Work /Item(s)

| | |
|---|---|
| Title | Tender for Supply and Installation of Cryogenfree Transport Set up with Accessories |
| Work Description | Tender for Supply and Installation of Cryogenfree Transport Set up with Accessories |
| Pre Qualification Details | Please refer Tender documents. |
| Independent External Monitor/Remarks | NA |

| | | | | | |
|--------------------------------|-----------|----------------------------------|-------------------------------------|------------------------------|----------------------|
| Tender Value in ₹ | NA | Product Category | Laboratory and scientific equipment | Sub category | NA |
| Contract Type | Tender | Bid Validity(Days) | 180 | Period Of Work(Days) | 30 |
| Location | IIT Ropar | Pincode | 140001 | Pre Bid Meeting Place | NA |
| Pre Bid Meeting Address | NA | Pre Bid Meeting Date | NA | Bid Opening Place | M Visvesvaraya Block |
| Should Allow NDA Tender | No | Allow Preferential Bidder | No | | |

Critical Dates

| | | | |
|--|----------------------|--|----------------------|
| Publish Date | 25-Aug-2021 05:00 PM | Bid Opening Date | 14-Sep-2021 03:30 PM |
| Document Download / Sale Start Date | 25-Aug-2021 05:00 PM | Document Download / Sale End Date | 14-Sep-2021 03:00 PM |
| Clarification Start Date | 25-Aug-2021 05:00 PM | Clarification End Date | 09-Sep-2021 11:00 AM |
| Bid Submission Start Date | 25-Aug-2021 05:00 PM | Bid Submission End Date | 14-Sep-2021 03:00 PM |

Tender Documents

| NIT Document | S.No | Document Name | Description | Document Size (in KB) |
|--------------|------|---------------|--------------------|---|
| | | 1 | Tendernotice_1.pdf | Tender for Supply and Installation of Cryogenfree Transport Set up with Accessories |

| Work Item Documents | S.No | Document Type | Document Name | Description | Document Size (in KB) |
|---------------------|------|------------------|---------------|---|-----------------------|
| | | 1 | BOQ | BOQ_677993.xls | BOQ FOR QUOTING RATES |
| | 2 | Tender Documents | UPLOAD.pdf | Tender for Supply and Installation of Cryogenfree Transport Set up with Accessories | 2704.51 |

Tender Inviting Authority

| | |
|----------------|---|
| Name | Deputy Registrar Store and Purchase M Visvesvaraya |
| Address | Deputy Registrar Store and Purchase M Visvesvaraya Block Indian Institute of Technology Ropar Rupnagar 140001 |



भारतीय प्रौद्योगिकी संस्थान रोपड़
INDIAN INSTITUTE OF TECHNOLOGY ROPAR

रूपनगर, पंजाब-140001/ Rupnagar, Punjab-140001
Ph. 01881-231285, 231283, e-mail: purchase@iitrpr.ac.in

File No. 1628-21/PH-10158/SERB-CRG+INST/GTE/PS/

Dated 25/08/2021

Indian Institute of Technology Ropar is in the process of purchasing following item(s) as per details as given as:-

| | |
|--|--|
| Details of the item | Cryogenfree Transport Set up with Accessories |
| Earnest Money Deposit to be submitted | N/A |
| Warranty | 3 years comprehensive onsite warranty |
| Delivery Schedule | As per tender document |

Tender Documents may be downloaded from Central Public Procurement Portal <http://eprocure.gov.in/eprocure/app>. Aspiring Bidders who have not enrolled / registered in e-procurement should enroll / register before participating through the website <http://eprocure.gov.in/eprocure/app>. The portal enrolment is free of cost. Bidders are advised to go through instructions provided at 'Instructions for online Bid Submission'.

Tenderers can access tender documents on the website (For searching in the NIC site, kindly go to Tender Search option and type 'IIT'. Thereafter, Click on "GO" button to view all IIT Ropar tenders). Select the appropriate tender and fill them with all relevant information and submit the completed tender document online on the website <http://eprocure.gov.in/eprocure/app> as per the schedule given in the next page.

No manual bids will be accepted. All quotation (both Technical and Financial should be submitted in the E-procurement portal).

(Registrar)

| SCHEDULE | |
|--|--|
| Name of Organization | Indian Institute of Technology Ropar |
| Tender Type (Open/Limited/EOI/Auction/Single/Global) | Global |
| Tender Category (Services/Goods/works) | Goods |
| Type/Form of Contract (Work/Supply//Service/Buy/Empanelment) | Supply |
| Product Category (Civil Works/Electrical Works/Fleet Management/ Computer Systems/Lab Equipment) | Lab Equipment |
| Date of Issue/Publishing | 25/08/2021 (17:00 Hrs) |
| Document Download/Sale Start Date | 25/08/2021 (17:00 Hrs) |
| Document Download/Sale End Date | 14/09/2021 (15:00 Hrs) |
| Last Date and Time for Uploading of Bids | 14/09/2021 (15:00 Hrs) |
| Date and Time of Opening of Technical Bids | 14/09/2021 (15:30 Hrs) |
| Tender Fee/EMD | Rs. <u> NIL </u> /- (For Tender Fee) Rs. <u> NIL </u> (For EMD) |
| | (To be paid through RTGS/NEFT. IIT Ropar Revenue Account Bank details are as under: |
| | Name of the Bank A/C : IIT Ropar Revenue Account |
| | SBI A/C No. : 37360100716 |
| | Name of the Bank : State Bank of India |
| | IFSC Code : SBIN0013181 |
| | MICR Code : 140002008 |
| | (This is mandatory that UTR Number is provided in the on- line quotation/bid. (Kindly refer to the UTR Column of the Declaration Sheet at Annexure-II) |
| No. of Covers (1/2/3/4) | 2 |
| Bid Validity days (180/120/90/60/30) | 180 days (From last date of opening of tender) |
| Address for Communication | Joint Registrar, Store & Purchase, M. Visvesvaraya Building, Indian Institute of Technology Ropar, Rupnagar – 140001 |
| Contact No. | 01881-231283,85 |
| Email Address | purchase@iitrpr.ac.in , drsp@iitrpr.ac.in |

Registrar

Instructions for Online Bid Submission:

As per the directives of Department of Expenditure, this tender document has been published on the Central Public Procurement Portal ([URL:http://eprocure.gov.in/eprocure/app](http://eprocure.gov.in/eprocure/app)). The bidders are required to submit soft copies of their bids electronically on the CPP Portal, using valid Digital Signature Certificates. The instructions given below are meant to assist the bidders in registering on the CPP Portal, prepare their bids in accordance with the requirements and submitting their bids online on the CPP Portal.

More information useful for submitting online bids on the CPP Portal may be obtained at:

<http://eprocure.gov.in/eprocure/app>

REGISTRATION

- 1) Bidders are required to enroll on the e-Procurement module of the Central Public Procurement Portal ([URL:http://eprocure.gov.in/eprocure/app](http://eprocure.gov.in/eprocure/app)) by clicking on the link "Click here to Enroll". Enrolment on the CPP Portal is free of charge.
- 2) As part of the enrolment process, the bidders will be required to choose a unique username and assign a password for their accounts.
- 3) Bidders are advised to register their valid email address and mobile numbers as part of the registration process. These would be used for any communication from the CPP Portal.
- 4) Upon enrolment, the bidders will be required to register their valid Digital Signature Certificate (Class II or Class III Certificates with signing key usage) issued by any Certifying Authority recognized by CCA India (e.g. Sify / TCS / nCode / eMudhra etc.), with their profile.
- 5) Only one valid DSC should be registered by a bidder. Please note that the bidders are responsible to ensure that they do not lend their DSCs to others which may lead to misuse.
- 6) Bidder then logs in to the site through the secured log-in by entering their userID / password and the password of the DSC / eToken.

SEARCHING FOR TENDER DOCUMENTS/

- 1) There are various search options built in the CPP Portal, to facilitate bidders to search active tenders by several parameters. These parameters could include Tender ID, organization name, location, date, value, etc. There is also an option of advanced search for tenders, wherein the bidders may combine a number of search parameters such as organization name, form of contract, location, date, other keywords etc. to search for a tender published on the CPP Portal.
- 2) Once the bidders have selected the tenders they are interested in, they may download the required documents / tender schedules. These tenders can be moved to the respective 'My Tenders' folder. This would enable the CPP Portal to intimate the

bidders through SMS / e-mail in case there is any corrigendum issued to the tender document.

- 3) The bidder should make a note of the unique Tender ID assigned to each tender, in case they want to obtain any clarification / help from the Helpdesk.

PREPARATION OF BIDS

- 1) Bidder should take into account any corrigendum published on the tender document before submitting their bids.
- 2) Please go through the tender advertisement and the tender document carefully to understand the documents required to be submitted as part of the bid. Please note the number of covers in which the bid documents have to be submitted, the number of documents - including the names and content of each of the document that need to be submitted. Any deviations from these may lead to rejection of the bid.
- 3) Bidder, in advance, should get ready the bid documents to be submitted as indicated in the tender document / schedule and generally, they can be in PDF / XLS / RAR / DWF formats. Bid documents may be scanned with 100 dpi with black and white option.
- 4) To avoid the time and effort required in uploading the same set of standard documents which are required to be submitted as a part of every bid, a provision of uploading such standard documents (e.g. PAN card copy, annual reports, auditor certificates etc.) has been provided to the bidders. Bidders can use “My Space” area available to them to upload such documents. These documents may be directly submitted from the “My Space” area while submitting a bid, and need not be uploaded again and again. This will lead to a reduction in the time required for bid submission process.

SUBMISSION OF BIDS

- 1) Bidder should log into the site well in advance for bid submission so that he/she upload the bid in time i.e. on or before the bid submission time. Bidder will be responsible for any delay due to other issues.
- 2) The bidder has to digitally sign the bid document and upload the required bid documents one by one as indicated in the tender document.
- 3) Bidder has to select the payment option as “on-line” to pay the tender fee / EMD as applicable and enter details of the instrument. Whenever, an EMD / Tender fee is sought, bidders need to pay the tender fee and EMD separately on-line through RTGS.
- 4) A standard BoQ format has been provided with the tender document to be filled by all the bidders. Bidders are requested to note that they should necessarily submit their financial bids in the format provided and no other format is acceptable. Bidders are required to download the BoQ file, open it and complete the white colored (unprotected) cells with their respective financial quotes and other details (such as name of the bidder). No other cells should be changed. Once the details have been completed, the bidder should save it and submit it online, without changing the filename. If the BoQ file is found to be modified by the bidder, the bid will be rejected.

- 5) The server time (which is displayed on the bidders' dashboard) will be considered as the standard time for referencing the deadlines for submission of the bids by the bidders, opening of bids etc. The bidders should follow this time during bid submission.
- 6) All the documents being submitted by the bidders would be encrypted using PKI encryption techniques to ensure the secrecy of the data. The data entered cannot be viewed by unauthorized persons until the time of bid opening. The confidentiality of the bids is maintained using the secured Socket Layer 128 bit encryption technology. Data storage encryption of sensitive fields is done.
- 7) The uploaded tender documents become readable only after the tender opening by the authorized bid openers.
- 8) Upon the successful and timely submission of bids, the portal will give a successful bid submission message & a bid summary will be displayed with the bid no. and the date & time of submission of the bid with all other relevant details.
- 9) Kindly add scanned PDF of all relevant documents in a single PDF file of compliance sheet.

ASSISTANCE TO BIDDERS

- 1) Any queries relating to the tender document and the terms and conditions contained therein should be addressed to the Tender Inviting Authority for a tender or the relevant contact person indicated in the tender.
- 2) Any queries relating to the process of online bid submission or queries relating to CPP Portal in general may be directed to the 24x7 CPP Portal Helpdesk. The contact number for the helpdesk is 1800 233 7315.

General Instructions to the Bidders

- 1) The tenders will be received online through portal <http://eprocure.gov.in/eprocure/app>. In the Technical Bids, the bidders are required to upload all the documents in .pdf format.
- 2) Possession of a Valid Class II/III Digital Signature Certificate (DSC) in the form of smart card/e-token in the company's name is a prerequisite for registration and participating in the bid submission activities through <https://eprocure.gov.in/eprocure/app>. Digital Signature Certificates can be obtained from the authorized certifying agencies, details of which are available in the web site <https://eprocure.gov.in/eprocure/app> under the link "Information about DSC".
- 3) Tenderer are advised to follow the instructions provided in the 'Instructions to the Tenderer the e-submission of the bids online through the Central Public Procurement Portal for e Procurement at <https://eprocure.gov.in/eprocure/app>.



TENDER DOCUMENT

TENDER FOR THE PURCHASE OF

Cryogenfree Transport Set up with Accessories

UNDER

TWO BID SYSTEM

No. 1628-21/PH-10158/SERB-CRG+INST/GTE/PS/

CHECK LIST

DULY FILLED CHECK LIST TO BE ATTACHED WITH THE TECHNICAL BID

| Sl. No. | Particulars | Check Mark |
|---------|---|------------|
| 1 | Whether EMD declaration attached? | Yes/ No |
| 2 | Whether technical specifications of the quoted equipment attached? | Yes/ No |
| 3 | Whether catalog of the equipment attached? | Yes/ No |
| 4 | In case of authorized agent/distributor whether certificate/ authorization letter for the same issued by the manufacturer attached? | Yes/ No |
| 5 | Whether tender document along with all Annexures duly signed & stamped by the authorized signatory attached? | Yes/ No |
| 6 | Whether affidavit duly attested by the Oath Commissioner/Executive Magistrate regarding non-black listing of supplier attached? | Yes/ No |
| 7 | Whether list of Institutes/Organizations where the quoted model of equipment supplied by the tenderer in India is attached? | Yes/ No |
| 8 | Whether split rates of each sub units are quoted? | Yes/ No |
| 9 | In case of foreign suppliers quoting directly, whether, the name of Indian agent mentioned? | Yes/ No |
| 10 | Whether compliance sheet duly filled in, signed & stamped attached? | Yes/ No |
| 11 | Whether warranty certificate duly filled in, signed & stamped attached? | Yes/ No |
| 12 | Whether Bank Details duly filled in, signed & stamped attached?k | Yes/ No |

List of Annexures

| Annexure | Particulars in annexures | Page No. |
|-----------------|---|-----------------|
| A | Technical Specifications | 12-17 |
| B | Format for Technical Compliance Sheet | 18 |
| C | Format for Manufacturer's Authorization Certificate | 19 |
| D | Format for Non-blacklisting Certificate | 20 |
| E | Format for User List | 21 |
| F | Format for Warranty Certificate | 22 |
| G | Bank Details of the Bidder | 23 |
| H | Declaration | 24 |
| I | Declaration | 25 |
| J | Declaration | 26 |
| K | Declaration | 27 |
| L | Declaration of County of Origin | 28 |



भारतीय प्रौद्योगिकी संस्थान रोपड़
INDIAN INSTITUTE OF TECHNOLOGY ROPAR
रूपनगर, पंजाब-140001/ Rupnagar, Punjab-140001
Ph. 01881-231285, 231283, e-mail: purchase@iitrpr.ac.in

No. 1628-21/PH-10158/SERB-CRG+INST/GTE/PS/

25/08/2021

Notice Inviting Quotation

Indian Institute of Technology Ropar invites online Bids (Technical bid and Commercial bid) from eligible and experienced OEM (Original Equipment Manufacturer) OR OEM Authorized Dealer for <Supply & Installation of Cryogenfree Transport Set up with Accessories> with (warranty period as stated at page #1 of this tender) on site comprehensive warranty from the date of receipt of the material as per terms & conditions specified in the tender document, which is available on CPP Portal <http://eprocure.gov.in/eprocure/app>-

| Sl. No. | Description | Quantity |
|---------|---|----------|
| 1 | Cryogenfree Transport Set up with Accessories (Detailed specifications are provided in Annexure - A) | 1 |

1. Schedule of Tender and EMD :

| | | |
|---|--|---------------------------|
| 1 | Last Date and Time for Uploading of Bids | 14.09.2021 up to 03:00 PM |
| 2 | Date and Time of Opening of Technical Bids | 14.09.2021 at 03:30 PM. |
| 3 | Earnest Money Deposit (EMD) | Nil |

A complete set of tender documents may be Download by prospective bidder free of cost from the website <http://eprocure.gov.in/eprocure/app>.

2. EMD:

The tenderer should submit an EMD declaration as per Annexure- K as per the Office Memorandum no. F-9/11/2020- PPD dt 12.11.2020 issued by Government of India.

3. Preparation of Bids:

- (a) Technical bid consisting of all technical details alongwith commercial terms and conditions and EMD Declaration; and
(b) Financial bid indicating item-wise price for the items mentioned in the technical bid.

The offer/bid should be submitted in two bid systems (i.e.) Technical bid and financial bid. The technical bid should consist of all technical details along with commercial terms and conditions. Financial bid should indicate item wise price for the items mentioned in the technical bid in the given format i.e BoQ_XXXX. The Technical bid and the financial bid should be submitted Online.

4. Submission of tender:

The tender has to be submitted on-line before the due date. The offers received after the due date and time will not be considered. No manual bids will be considered.

5. Bidding:

- a). Items of indigenous nature or quoted in INR, more than 1 authorized representative may participate in the same tender and submit their bids on behalf of their OEM/Principal/Manufacturer if the OEM permits more than one authorized bidder in such case as per their policy.
b). In cases of agents quoting in offshore procurements, on behalf of their principal manufacturers, one agent cannot represent two manufacturers or quote on their behalf in a particular tender enquiry. One manufacturer can also authorize only one agent/dealer
c). The letter of authority should be on the letterhead of the manufacturer and should be signed by a person competent and having the power of attorney to bind the manufacturer. The same should be included by the bidder in its techno-commercial unpriced bid.
All offers other than those from the Principal/OEM should be supported by an authority letter from the manufacturer authorizing the supplier to tender on their behalf. In case of manufacturer a certificate or a copy thereof to the effect that the bidder is a manufacturer of the equipment must be accompanied with the technical bid prepared as per 'Annexure - C'.

6. Opening of Bids:

The online bid will be opened by a committee duly constituted for this purpose. Online bids (complete in all respect) received will be opened as mentioned in the "Schedule" in presence of bidders representative if available. Only one representative will be allowed to participate in the tender opening. Bid received without declaration will be rejected straight way. The technical bid will be opened online first and it will be examined by a technical committee (as per specification and requirement). The financial offer/bid will be opened only for the offer/bid which technically meets all requirements as per the specification, and will be opened in the presence of the vendor's representatives subsequently for further evaluation. The bidders if interested may participate on the tender opening Date and Time. The bidder should produce authorization letter from their company to participate in the tender opening

7. Price Bid:

Price bid should be submitted in given BOQ_XXXX.xls format online.

The price should be quoted in net per unit (after breakup) and must include all packing and delivery charges. Price quoted should be in Indian Rupees, free delivery at IIT Ropar Campus at site (DDP/FOR).

Further, depending on the nature of the goods, there may be cost elements towards installation and commissioning, operator's training, and so on. Normally, it may be included in the equipment cost but if it is quoted separately, the same will be added in the item price for the determination of ranking of the bidders. The offer/bid should be exclusive of taxes and duties, which will be paid by the purchaser as applicable. However, the percentage of taxes & duties shall be clearly indicated. Necessary certificate will be issued on demand. The Buyer/PFC will have the right to award contracts to different Bidders for being lowest in particular items.

For ranking of offers, price of complete scope of supply as detailed in technical specifications, the procuring authority/Purchaser may decide as follows for comparison of price bid –

- (i) All items of the bid which are mandatorily required to meet the tendered specifications of the item/system
(ii) If a bidder has put certain items/modules which are required to meet the tendered specifications in the 'optional' part of the bid, then such optional items shall also be included for the purpose of price comparison
(iii) On the other hand, if a bidder has inadvertently included any item/module in its main price bid which is not required as per tender specifications, then the price of such item/module shall be excluded from the price comparison provided that the price for the said item/module is clearly reflected separately in the bid
(iv) Anything asked as 'optional' in our specs is not to be included for overall comparison

Non-conformities between Figures and words: Sometimes, non-conformities/errors are also observed in responsive tenders between the quoted prices in figures and in words. This situation normally does not arise in case of e-Procurement. This should be taken care of in the manner indicated below:

- (i) If there is a discrepancy between the unit price and the total price that is obtained by multiplying the unit price and quantity, the unit price will prevail and the total price will be corrected.
(ii) If there is an error in a total corresponding to the addition or subtraction of sub-totals, the sub-totals shall prevail and the total shall be corrected;
(iii) If there is a discrepancy between words and figures, the amount in words will prevail for calculation of price.

8. Taxes:

Suppliers shall be entirely responsible for all taxes, duties, license fees, octroi, road permits, etc., incurred until delivery of the contracted Goods to the Purchaser. However, GST etc, in respect of the transaction between the Purchaser and the Supplier shall be payable extra, if so stipulated in the order. For research purpose(s) ONLY, 5% GST will be applicable with concessional GST Certificate.

9. Duties:

IIT Ropar is exempted from paying custom duty under notification No.51/96 (partially or full) and necessary "Custom Duty Exemption Certificate" can be issued after providing following information and Custom Duty Exemption Certificate will be issued to the shipment in the name of the Institute, (no certificate will be issued to third party): The procured product should be used for teaching, scientific and research work only.

- a) Shipping details i.e. Master Airway Bill No. and House Airway No. (if exists)
- b) Forwarder details i.e. Name, Contact No., etc.

IIT Ropar is partially exempted from paying GST and necessary GST Exemption Certificate will be provided for which following information are required.

- b) Quotation with details of Basic Price, Rate, Tax & Amount on which ED is applicable
- c) Supply Order Copy
- d) Proforma-Invoice Copy.

10. Rates Comparison:

Bidders are requested to send their rates on FOR, IIT Ropar basis in case of indigenous items and on FOB/FCA basis in case of imported items with separately mentioning CIF/CIP charges. Rates comparison will be made on Net Price (Including Freight/ Insurance/Installation/Taxes/Duties etc.) if offers are received both for indigenous and import items. If offers are received only for import items then comparison will be made on FOB/FCA basis. Rates of foreign currency will be taken from RBI website as on the opening of the technical bids for the purpose of conversion in Indian rupees.

11. Spares:

The spare parts/wear & tear consumables, if any, required for trouble free operation of equipment to be quoted separately giving the full nomenclature, rate, quantity and shelf life of each item.

12. Indigenous items:

The items which can/are to be provided indigenously may be listed separately.

13. Parts of Equipments:

Where the equipment is composed of several subunits/components, the rate should be quoted for each subunit/component. The Institute reserves the right to increase or decrease the number of subunits/components and number of equipment according to its requirements. The rates in ambiguous terms will render the quotation liable to rejection. The words "Not quoting" should be clearly written against any item of equipment for which the tenderer is not quoting.

14. Payment Terms:

Payment will be made to the supplier through following modes.

a). Indigenous goods:

- i) NEFT/Cheque/Demand Draft : 90% payment will be made within 30 days from the date of receipt of material at IIT Ropar and balance 10% after successful installation of the equipment and on the submission of performance bank guarantee for 3% of order value, valid for warranty period + 3 months or
- ii) NEFT/Cheque/Demand Draft : 100% payment will be made within 30 days from the date of receipt of material at IIT Ropar and successful installation of the equipment.

b). Imported goods:

Letter of credit/Telegraphic Transfer/Sight Draft – 90% payment will be made through LC/TT/SD and balance 10% after successful installation of the equipment and submission of performance bank guarantee for 3% of order value, either by the principal company or by their Indian agent valid for warranty period + 3 months.

Bank charges occurred outside India will be borne by the beneficiary.

15. Acceptance of Terms & Conditions:

Bidders must confirm the acceptance of all the terms and conditions of this NIQ. Any non-acceptance or deviations from the terms and conditions must be clearly mentioned. However, tenderers must note carefully that any conditional offer or any deviation from the terms and conditions of this NIQ may render the quotation liable for rejection.

16. Service Manual/Circuit Diagram

It is specifically required that the bidders will supply all the operating & service manuals and circuit diagrams alongwith the equipment.

17. Power Supply:

The equipment should be quoted only for 220 volts and 50Hz electricity supply. The extra requirement of line voltage, current rating etc. and the optimum climate and environment required for the equipment must be stated precisely. Voltage stabilizers/isolation transformers/CVT/UPS etc., as may be required shall be listed separately. The full technical specifications and literature in respect of the voltage stabilizer etc., must be furnished.

18. Guarantee/Warranty:

Duly signed and stamped certificate of at least 3 years comprehensive onsite warranty as per Annexure-F should be attached with the technical bid. Successful firm will be required to agree for payment of penalty for exceeding permissible downtime during Guarantee / Warranty period.

19. Country of origin:

Country of origin of the quoted item should be mentioned in the offer in case of imported item.

As per Ministry of Finance, Deptt. of Expenditure, Public Procurement Division Order (Public Procurement No.1) issued from file No.6/18/2019-PPD dated 23rd July, 2020 regarding Restrictions under Rule 144 (xi) of the General Financial Rules (GFRs) 2017, it is directed that any bidder from a country which shares a land border with India will be eligible to bid in any procurement whether of goods, services (including consultancy services and non-consultancy services) or works (including turnkey projects) only if the bidder is registered with the Competent Authority i.e. the Deptt. for Promotion of Industry and Internal Trade (DPIIT). The said order will not apply to bidders from those countries (even sharing a land border with India) to which the Government of India has extended lines of credit or in which the Government of India is engaged in development projects (updated lists of the countries are given in the Ministry of External Affairs)

"Bidder" (including the term 'tenderer', 'consultant' or 'service provider' in certain contexts) means any person or firm or company, including any member of a consortium or joint venture (that is an association of several persons, or firms or companies), every artificial juridical person not falling in any of the descriptions of bidders stated hereinbefore, including any agency branch or office controlled by such person, participated in a procurement process.

"Bidders from a country which shares a land border with India" for the purpose of this Order means:

- i. An entity incorporated, established or registered in such a country; or
- ii. A subsidiary of an entity incorporated, established or registered in such a country; or
- iii. An entity substantially controlled through entities incorporated, established or registered in such a country; or
- iv. An entity whose beneficial owner is situated in such a country; or
- v. An Indian (or other) agent of such an entity; or
- vi. A natural person who is the citizen of such a country; or
- vii. A consortium or joint venture where any member of the consortium or joint venture falls under any of the above

The beneficial owner for the purpose of above will be as under: -

1. In case of a company or Limited Liability Partnership, the beneficial owner is the natural person (s), who, whether acting alone or together, or through one or more juridical person, has a controlling ownership interest or who exercise control through other means.

Explanation

a. "Controlling ownership interest" means ownership of or entitlement to more than twenty-five per cent of share or capital or profit of the company;
b. "Control" shall include the right to appoint majority of the directors or to control the management of policy decisions including by virtue of their shareholding or management rights or shareholders agreements or voting agreements;

2. In case of a partnership firm, the beneficial owner is the natural person (s) who, whether acting alone or together, or through one or more juridical person, has ownership of entitlement to more than fifteen percent of capital or profits of the partnership;

3. In case of an unincorporated association or body of individuals, the beneficial owner is the natural person (s), who, whether acting alone or together, or through one or more juridical person, has ownership of or entitlement to more than fifteen percent of the property or capital or profits of such association or body of individuals;

4. Where no natural person is identified under (1) or (2) or (3) above, the beneficial owner is the relevant natural person who holds the position of senior managing official;

5. In case of a trust, the identification of beneficial owner(s) shall include identification of the author of the trust, the trustee, the beneficiaries with fifteen percent or more interest in the trust and any other natural person exercising ultimate effective control over the trust through a chain of control or ownership.

An agent is a person employed to do any act for another, or to represent another in dealings with the third person.

For Works contracts, including Turnkey contracts, the successful bidder shall not be allowed to subcontract works to any contractor from a country which shares a land border with India unless such contractor is registered with the Competent Authority.

A certificate shall be submitted by bidders in the tender documents regarding their compliance with the said order. If the certificate submitted by a bidder whose bid is accepted is found to be false, this would be a ground for immediate termination and further legal action in accordance with law. Annexure I (For Goods/ Services contracts)/ Annexure J (For Works contracts, including Turnkey contracts)

20. Service Facility:

Bidder should mention about the service set up in India and how capable they are to provide after sales services.

21. Training:

If required, should be included in your offer without any extra cost.

22. Banker's details:

Name and address of the banker of your company should be mentioned.

23. Reference of supply:

Name and contact details of the premier educational Institutes where the quoted equipment has been installed in India should be attached as per Annexure-E. Copies of at least two purchase orders may be attached (If possible). IIT Ropar reserves the right to inspect the equipment for its actual performance in any of the listed Institute.

24. Arbitration

In the event of failure to carry out the contractual obligations, within the stipulated period or extended period and determination of the contract for any reason, violation of warranties etc. the IIT Ropar shall have the right to carry out the unfinished obligation at the exclusive cost and risk of the bidder/firm, after due notice and the difference so accrued shall be recoverable from the bidder/firm.

23.2 The provision of the Arbitration and Conciliation Act, 1996 or as at the relevant time and of rules framed there under and any statutory modifications thereof shall be deemed to apply and be incorporated in this agreement.

23.3 Upon every or any such reference, the cost of any incidentals to the reference and award(s) respectively shall be at the reasonable discretion of the Arbitrators or in the event of their not agreeing, of the Umpire appointed by them who may determine the amount thereof or direct the same to be fixed as between solicitors and client or as between parties and shall direct by whom and in what manner the same shall be borne and paid.

23.4 Panel of arbitrators will be provided by IIT Ropar out of which the bidder will have to select one.

23.5 The bidder shall have no objection if the officer who has dealt with the case at any stage is nominated as an arbitrator. Further, that one of the arbitrator's shall be Accounts Expert.

23.6 In case of vacancy being caused due to resignation, death or incapacity of the arbitrator(s) to function as such, the same shall be provided in the aforesaid manner and the new arbitrator(s) shall proceed from the stage at which vacancy is caused.

25. Jurisdiction:

The Courts of Ropar alone will have the jurisdiction to try any matter, dispute or difference between the parties arising out of this tender/contract. It is specifically agreed that no Court outside and other than Ropar court shall have jurisdiction in the matter.

26. Force Majeure:

Any failure of omission or commission to carry out the provision of this contract by the supplier shall not give rise to any claim by one party, one against the other, if such failure of omission or commission arises from an act of God; which shall include all acts of natural calamities such as fire, flood, earthquake, hurricane, or may pestilence or from civil strikes, compliance with any statute and/or regulations of the Government, lockouts and strikes, riots, embargoes or from any political or other reason beyond the supplier's control including war (whether declared or not) civil war or stage of insurrection, provided that notice of the occurrence of any event by either party to the other shall be given within two weeks from the date of occurrence of such an event which could be attributed to Force Majeure conditions.

27. Risk & Cost

In the event of failure to carry out the contractual obligations, within the stipulated period or extended period and determination of the contract for any reason, violation of warranties etc. the IIT Ropar shall have the right to carry out the unfinished obligation at the exclusive cost and risk of the bidder/firm, after due notice and the difference so accrued shall be recoverable from the bidder/firm.

28. The material found defective upon opening by the supplier representative in presence of Central stores personnel/indenter of IIT Ropar or not as per tendered specifications will be lifted back at the cost and risk of the supplier. The material lying in the IIT Ropar premises would be at supplier's risk and cost.

29. Liquidated Damages:

In case the firm fails to execute the supply as per the purchase order in whole or in part as per the terms and conditions of PO, IIT Ropar can impose the penalty @1% per week of the undelivered stores, subject to a maximum of 10%. It will also be open to the institute to procure the required item(s) from any other source at the risk and expense of the firm.

30. Relocation:

The supplier has to stand guaranteed for the relocation of supplied equipment once the permanent campus of IIT Ropar gets ready for operation. Transportation of the equipment will be provided by the Institute.

Note: The Director, IIT Ropar reserves the right to accept/reject any or all tenders without assigning any reasons thereof and also to reject the material if the same is not found conforming to the specifications, with further right to affect risk and cost of the purchases.

Registrar

Requirement for High-field and Low-temperature based Measurement System

A cryogen-free transport measurement set up with temperature of operation in the range of 1.8 K to 300 K and magnetic fields up to 9 Tesla. System should be fully cryogen free, i.e. no requirements of liquid Helium. The system should be fully automated and computer controlled with latest technology for instrument interfacing and control in whole range of measurements. The system should NOT have any manual control throughout the measurements range with provision for addition of user specific programming cycles for measurements and control.

1. Base System:

- a) System should be fully cryogen-free i.e. no requirements of liquid Helium. Both the superconducting magnet and the sample space should be cooled by:
 - (i) A single pulse tube (two stage) closed cycle refrigerator (CCR) of appropriate cooling capacity of cold head and compressor
 - (ii) The Variable Temperature Insert (VTI) should be a continuous flow closed circuit design where the cooling of the VTI and sample is provided by the circulation of Helium.
- b) Any Liquid Helium and/or cold Helium gas flow into sample chamber or to any other parts within the system, and all low temperature operations must be fully automated and user friendly. The system should not have any manual control in the entire operation of the cryo-system. Vendor should describe in detail the temperature control mechanism, stability, reproducibility and accuracy.
- c) Sample space of at least 25 mm or better. The bore of the magnet should be at least 30mm or better in diameter.
- d) Variable Temperature Insert (VTI) with appropriate heat shields and sample holder should have an exchange gas option, for rapid cool down and appropriate temperature stability and control over the entire temperature range of operation, viz.: 1.8 to 300 K or better. Provide supporting data from similar installed systems and factory tests. Appropriate gas handling system for He gas, with flow and pressure regulators, filters for flushing and operation of the cryogen free VTI should be there.
- e) Calibrated temperature (Cernox) sensors mounted on the sample position and on the VTI heat exchanger for high precision temperature measurements in the presence or absence of magnetic field. Additional system diagnostic temperature sensors on cryocooler 1st and 2nd stages, 1.5 K pot, persistent switch (if desirable) and magnet.

- f) Magnet control software monitors the temperature of the magnet and cryostat at least at 3 locations to ensure proper operation of the magnet system and avoid time consuming quenches.
- g) System must have a suitable oil free dry pump (preferably Pfeiffer turbo molecular pumps) backed by Pfeiffer /Edwards dry two stage backing pumps) and appropriate vacuum gauges for monitoring the cryosystem and the sample environment. The vacuum pump must be able to pump out the sample chamber to less than 10^{-4} Torr in under 15 minutes. Details of the pumps and vacuum fittings must be provided.
- h) The sample space should be hermetically sealed from the VTI circuit in order to prevent accidental contamination of the circuit. The sample space should have sufficient size to accommodate all measurement options offered by the supplier whether purchased simultaneously with the system, or at a later date as an upgrade. The sample space should be filled with static exchange gas or evacuated depending on the operating temperature range, and if experiments in vacuum are required.
- i) The top loading sample insertion system should consist of Air Lock and Gate valve enabling the user to insert the sample while the Sample Chamber is still cold at ~ 2 to 10 K.
- a) Appropriate water-cooled helium compressor with full charge of high purity Helium gas, with at least 20 m flexible SS lines. Electrical and cooling water requirements for the compressor must be specified. Maintenance interval for the compressor must be at least 30,000 hours and for the cold head must be at least 20,000 hours. Preferable makes: Sumitomo or CryoMech.
- j) The entire system must be modular with an option to field upgrade to He-3 refrigerator in future with base temperature of ~ 300 mK.
- k) The cold head, motor head and the VTI should be electrically isolated. The vibration level at the sample space should be clearly mentioned.

I. Superconducting magnet:

- a) Longitudinal magnetic field of ± 9 T or higher should be available. (Quote optionally for 12 T)
- b) Sweep rate: 0.5 – 50 Oe/sec or larger for 9 T magnet.
- c) Field Homogeneity $\pm 0.1\%$ over 1 cm^3 volume.
- d) Field stability: 100 ppm/hr or better.
- e) Thermometer directly on the magnet. Automatic discharge of the magnet if the cryocooler system fails due to any exigency, eg: chiller water failure, etc.

- f) Appropriate four-quadrant magnet power supply for 9 T (12 T option) magnet with ability to continuously sweep the magnetic field to positive and negative values through zero. The Bi-polar power supply for the magnet with over voltage protection and indication.
- g) There should be no overshoot in the field or the tolerable overshoot in “No Overshoot” mode should be specified for various field strength. Please also specify the approach mode i.e. linear, oscillating or no-overshoot.
- h) Magnet must be adequately protected from quenches, and the cryosystem should be designed to handle any magnet quench, with minimum turnover time.
- i) Time to Full Field: 45 minutes or faster (vendor must specify and provide supporting info)
- j) Time to cool down the magnet from initial starting must be efficient (≤ 42 Hrs.). Magnet cool down data should be provided.

II. Temperature control:

- a) The VTI should be a closed circuit, continuous flow system and should be able to maintain stable temperature for indefinitely long time. The temperature range should be 1.8K or lower to 300 K or higher.
- b) The system should enable cooling of samples from highest temperature to the lowest at the highest specified cooling rate at any given magnetic field (up to 9 T/12 T) without affecting the system performance or heating of magnet or resulting in a magnet quench. System should be stable when sample heating or temperature control is ON.
- c) Temperature range of 1.8K to 300 K, with 50 milli Kelvin or better, stability and accuracy, with and without magnetic field.
- d) Fully automated cooling and warm up operation in the temperature range of 1.8 K to 300 K. Time required for cooling down from room temperature to stable 5K or warm up from 5 K to room temperature should be preferably not be more than 150 minutes. Vendor must provide supporting data.
- e) System must have efficient temperature control system enabling user to perform reliable and continuous, fully automated, low temperature operation. Measurement at $T < 4.2$ K should be possible for a long time (> 1 hour) without compromising the parameters mentioned above (temperature should be stable while ramping the magnetic field and for input of the sample current)
- f) Accuracy: $\pm 1\%$ and minimum sweeping rate 0.01K/min and maximum sweep rate of at least 5k/min while ramping the Temp and 2k/min while cooling irrespective of the magnitude of applied magnetic field. System with higher sweep rate is preferred.

Single axis Sample Rotator:

- a) Modular mechanical single axis sample rotator to rotate sample both in plane and from in-plane to out of plane magnetic field. This should work for sample in exchange gas and be operated via computer control and compatible with VTI.
- b) This rotator should have easily mountable and dismountable PCB sample loading platform with gold pads for electrical connections. It should have thermometer to closely monitor sample temperature. 20 pin or better chip carrier (such as LCC 20 as a chip holder) along with the suitable loading platform. (20 pcs of non-magnetic chip carriers should be supplied with the system).
- c) 10 numbers (or higher) twisted pair constantan/copper wires down to the sample holder (sample loading platform for LCC 20 or any other configuration). The wires should be connected with the loading platform and should have the rotation of 0° to 180° (-90° to 90°) or better. The wires should have suitable connectors terminating at the top to (preferably Fischer connector). Additional blank port at the top of the probe, with access to sample for additional wiring in future, if necessary. Moreover, in the midway of the wiring of the sample connector, there should be a break out switch (wire starting from the top terminating into a male connector and the rest of the wiring starting from a female connector going to the LCC20 or other suitable chip carrier platform) such that user could include some electronics components such as filters for future use.
- d) A thermometer must be in direct contact with the sample mount so as to enable the accurate determination of sample temperature. A Cernox or equivalent thermometer must be in close vicinity (body of the sample manipulator close to the sample mount) so as to enable the accurate determination of sample temperature.
- e) Angle of rotation should be in the range from 0° to 180° (-90° to 90°) or better, with a well-defined angular step size of less than or equal to 0.1° . Vendor must specify respective values.
- f) Rotator must be usable in the entire temperature range of 1.8 to 400 K and support resistivity measurements

At least 20 sets of spare nonmagnetic sample holder (for Horizontal sample rotator) should be included.

All the electrical components including compressor should be compatible with the Indian Standard system. The Indian standard system is $240V \pm 5\%$ for single phase and $415V \pm 5\%$ for three phase connection with frequency 50 Hz.

Additional items to be included:

1. Tool kit along with spares for all consumables, viz: 25 ml GE varnish, Apiezon N grease; Stycast, Molex pins (M & F); manganin, copper, cupro-nickel cable spools, 10 spare

connectors for midway breakout switch, spare pins and tools to remove and install the pins, sets of O-rings, retainer rings, clamps; sets of SS304, Brass, ETP Copernuts washers, bolts should be included with the system.

The instrument must carry a comprehensive warranty of 3 year from the date of installation.

Data acquisition and analysis:

- a) Licensed windows based operating software with workstation compatible with the measurement options. The computer should have 16 GB DDR4 RAM and i7 processor with windows 10 operating software and 1 TB SSD Hard Drive. LabView license should be provided in case the measurement software is based on Labview.

Optional Items:

2. Optionally quote for 12 T magnet.
3. Variable temperature He-3 insert capable of reaching base temperature of ~300 mK (without load) and maximum sample base temperature of 300 K and compatible with the above pulse tube-based cryogen free and superconducting system.
4. A measurement probe with a fixed sample position and two semi rigid high frequency coax lines (UT85 SS-SS frequency up to 40 GHz) with K type connector on one end and left loose at sample end. The loose end would be at height suitable such that it can be coupled with the sample at later stage.
5. Option to measure the sample in vacuum.
6. Double axis rotator which would allow the platform to rotate around the horizontal axis and the sample rotates on the platform around the center of the platform. The angular range is 180 degrees on both axis of rotation. The rotator should have a computer control. The double axis rotator must be compatible with the provided sample chip.
7. Ultra-Low Field to reduce the field in the range of 20 to 30 mT with step of 1microTesla.

4. Other Details:

- 1) Vendor must have a factory trained service support engineer in India. The contact details for the service support engineer for the after sales service should be mentioned. Also, all third party items such as pumps, gauges, power supplies, other electronic instruments and units, etc. should have service center in India. In case the vendor does not have any factory

trained local service support engineer based in India, the warranty of the system should be for 5 years.

- 2) Vendor must provide the detail specifications in the compliance statement with respect to each technical specification in the tender document duly supported by the manufacturer's literature and published papers.
Mere saying yes to compliance sheet without supporting data and measurement will not be accepted and would lead to the disqualification of the vendor.
- 3) Vendor must provide customer satisfaction letter/certificates from Indian Institutes/Labs where similar measurement systems are installed and working satisfactory, in the last 5years, and from foreign labs in the last 3 years. Installation reports are not a valid document for the said purpose. Those users will be contacted for system feedback.
- 4) List of similar items installed during last five years in institutes like IITs/NISER/IISERs/IISC/TIFR/SINP/RRCAT/IOP/SN Bose and any DRDO labs, India with contact person name, address and phone number, email id must be specified.
- 5) Proper drawings to be submitted for approval in case P.O. is received and before manufacturing commences.
- 6) Warranty: Minimum of 3 years from the date of installation. Comprehensive Annual Maintenance Contract, after the warranty period, must be clearly mentioned in bid.
- 7) All pumps gauges and control units must have service centers in India and suitable warranty and comprehensive AMC for the duration as specified in the purchase order should be included.
- 8) Pre-installation site preparation requirements to be included and specified along with the bid.
- 9) Technical evaluation by the institute may include demonstration to verify functionalities and capabilities of the system quoted. Institute may ask the companies to access the instruments remotely for checking the claimed functionalities. Institute can also ask for sample measurement, if required.
- 10) It should be Delivered duty paid (DDP) IIT Ropar. IIT Ropar would provide the custom duty exemption certificate.
- 11) The instrument should be shifted and re-installed in the main campus of the IIT Ropar in case the system gets installed in the transit campus of IIT Ropar, without any extra charge. The company should arrange for the insurance during the transport.

FORMAT FOR MANUFACTURER'S AUTHORISATION CERTIFICATE

To,
The Registrar
Indian Institute of Technology Ropar
Nangal Road, Rupnagar-140001

Sub. : Tender for “ _____ ”.

Dear Sir,

We manufacture of original equipment at (.....address of factory.....)
do hereby authorize M/s (Name and address of Agent) to submit a bid, negotiate and receive the order format against
your tender enquiry.

M/s. is authorized to bid and conclude the contract in regard to this business.

We hereby extend our full guarantee and warranty as per clause conditions NIQ for the
goods and services offered by the above firm.

Yours faithfully,

[Name & Signature]

For and on behalf of M/s. _____ [Name of manufacturer]

Note:

1. Items of indigenous nature or quoted in INR, more than 1 authorized representative may participate in the same tender and submit their bids on behalf of their OEM/Principal/Manufacturer if the OEM permits more than one authorized bidder in such case as per their policy.
2. In cases of agents quoting in offshore procurements, on behalf of their principal manufacturers, one agent cannot represent two manufacturers or quote on their behalf in a particular tender enquiry. One manufacturer can also authorize only one agent/dealer
3. The letter of authority should be on the letterhead of the manufacturer and should be signed by a person competent and having the power of attorney to bind the manufacturer. The same should be included by the bidder in its techno-commercial unpriced bid.

FORMAT FOR NON BLACKLISTING OF SUPPLIER

I/ We _____Manufacturer/partner/Authorized Distributor/Agent (strike out which is not applicable) of (Supplier) _____ do hereby declare and solemnly affirm that the individual/firm/company is not black-listed by the Union/State Government/Autonomous body.

Deponent
Address _____

I/ We hereby solemnly declare and affirm that the above declaration is true and correct to the best of my knowledge and belief. No part of it is false and nothing has been concealed.

Deponent
Dated: _____

(Note: To be furnished on non-judicial stamp paper duly attested by the Oath Commissioner.)

CERTIFICATE OF WARRANTY

i). I/We certify that the warranty shall be for a period of at least 3 years comprehensive onsite warranty starting from the date of satisfactory installation, commissioning and handing over of the equipment and of the works conducted therewith covered under the supply order in working order. During the warranty period, I/we shall provide free “after sale service” and the replacement of any part(s) of the equipment or rectification of defects of work of the equipment will be free of cost. The replacement of the parts shall be arranged by us, at our own cost and responsibility. We undertake that the above warranty shall begin only from the date of satisfactory and faultless functioning of the equipment for 60 days at IIT Ropar premises. The benefit of change in dates of the warranty period shall be in the interest of the use/your organization.

ii). During the warranty period, we shall provide at least 3 preventive maintenance visits.

iii). Uptime Guarantee: During the warranty period, we will be responsible to maintain the equipment in good working conditions for a period 328 days (i.e. 90% uptime) in a block of 365 days.

a). All complaints will be attended by us within 2 weeks of receipt of the complaint in our office.

b). In case there is delay of more than 2 weeks in attending to a complaint from our side then you can count the number of days in excess of the permissible response time in the downtime. The above said response time of 2 weeks for attending to a complaint by us will not be counted in the downtime.

c). Penalty: We shall pay a penalty equivalent to 0.1 % of the FOB value of the equipment for every week or part thereof delay in rectifying the defect.

Note: The right to accept the reason (s) for delay and consider reduction or waive off the penalty for the same shall be at the sole discretion of Director, IIT Ropar

iv. We certify that the equipment being/quoted is the latest model and that spares for the equipment will be available for a period of at least 10 years and we also guarantee that we will keep the organization informed of any update of the equipment over a period of 10 years.

v. We guarantee that in case we fail to carry out the maintenance within the stipulated period, IIT Ropar reserves the right to get the maintenance work carried out at our risk, cost and responsibility after informing us. All the expenses including excess payment for repairs/maintenance shall be adjusted against the Performance Bank Guarantee. In case the expenses exceed the amount of Performance Bank Guarantee, the same shall be recoverable from us with/without interest in accordance with the circumstances.

vi. We shall try to repair the equipment at IIT Ropar premises itself. However, the equipment will be taken to our site on our own expenses in case it is not possible to repair the same at IIT Ropar. We shall take the entire responsibility for the safe custody and transportation of the equipment taken out for repairs till the equipment is rehabilitated to the IIT Ropar after repair Any loss of equipment or its accessories under its charge on account of theft, fire or any other reasons shall be at our sole risk and responsibility which will be compensated to IIT Ropar for such losses at the FOB/CIF value for the damaged/lost equipment/part, including accessories.

vii. We undertake to perform calibration after every major repair/breakdown/taking the equipment for repair out of IIT Ropar premises.

viii. In case of extended warrantee, we undertake to carry out annual calibration of the equipment.

ix. We guarantee that we will supply spare parts if and when required on agreed basis for an agreed price. The agreed basis could be an agreed discount on the published catalogue price.

x. We guarantee to the effect that before going out of production of spare parts, we will give adequate advance notice to you so that you may undertake to procure the balance of the life time requirements of spare parts.

xi. We guarantee the entire unit against defects of manufacture, workmanship and poor quality of components.

ON THE COMPANY/FIRM LETTER HEAD

To,
The Registrar,
Indian Institute of Technology Ropar,
Punjab-140001

SUBJECT: Submission of RTGS/NEFT details

Sir,

As per your instructions, the detail of RTGS/NEFT in respect of M/s _____ is as follows:

| | |
|--------------------------|--|
| BENEFICIARY NAME :- | |
| BANK NAME:- | |
| BRANCH NAME:- | |
| BRANCH CODE:- | |
| BANK ADDRESS:- | |
| TYPE OF A/C.:- | |
| BANK A/C.NO.:- | |
| IFS CODE:- | |
| MICR NO:- | |
| PAN OF BENEFICIARY:- | |
| SERVICE TAX NO.:- | |
| TIN NO.:- | |
| NAME OF CONTACT PERSON:- | |
| TELEPHONE/MOBILE NO.:- | |
| EMAIL ID:- | |

Certified that above mentioned details are true and correct.

For M/s _____

(Authorised signatory)

FOR BANK USE ONLY

Verified the above mentioned detail and signature as per the records maintained by _____
(bank name).

Signature (with bank seal)

Code _____

DECLARATION SHEET

We, _____ hereby certify that all the information and data furnished by our organization with regard to this tender specification are true and complete to the best of our knowledge. I have gone through the specification, conditions and stipulations in details and agree to comply with the requirements and intent of specification. This is certified that our organization has been authorized (Copy attached) by the OEM to participate in Tender. We further certified that our organization meets all the conditions of eligibility criteria laid down in this tender document. Moreover, OEM has agreed to support on regular basis with technology / product updates and extend support for the warranty. The prices quoted in the financial bids are subsidized due to academic discount given to IIT Ropar

| We, further specifically certify that our organization has not been Black Listed/De Listed or put to any Holiday by any Institutional Agency/ Govt. Department/ Public Sector Undertaking. | | NAME & ADDRESS OF THE Vendor/ Manufacturer / Agent |
|---|---------------------|---|
| 1 | Phone | |
| 2 | Fax | |
| 3 | E-mail | |
| 4 | Contact Person Name | |
| 5 | Mobile Number | |
| 6 | GST Number | |
| 7 | PAN Number | |

(Signature of the Tenderer)

Name: _____

Seal of the Company

<On Organization Letter Head>

ANNEXURE-I
(For Goods/ Services Contracts)

<CERTIFICATE>

Tender No. :-

Date:-

I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and hereby certify that this bidder is not from such a country.

OR (whichever is applicable)

I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and hereby certify that this bidder is from _____ (Name of Country) and has been registered with the Competent Authority. I also certify that this bidder fulfills all the requirements in this regard and is eligible to be considered.

(Copy/ evidence of valid registration by the Competent Authority is to be attached)

Signature of Bidder/ Agent
Name: _____
Designation: _____
Organization Name: _____
Contact No. : _____

<On Organization Letter Head>

ANNEXURE-J
(For Works Contracts, including Turnkey contracts)

<CERTIFICATE>

Tender No. :-

Date:-

I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and on sub-contracting to contractors from such countries and hereby certify that this bidder is not from such a country and will not sub-contract any work to a contractor from such countries unless such contractor is registered with the Competent Authority.

OR (whichever is applicable)

I have read the clause regarding restrictions on procurement from a bidder of a country which shares a land border with India and on sub-contracting to contractors from such countries and hereby certify that this bidder is from _____(Name of Country) and has been registered with the Competent Authority and will not sub-contract any work to a contractor from such countries unless such contractor is registered with the Competent Authority. I also certify that this bidder fulfills all the requirements in this regard and is eligible to be considered.

(Copy/ evidence of valid registration by the Competent Authority is to be attached)

Signature of Bidder/ Agent

Name: _____

Designation: _____

Organization Name: _____

Contact No. : _____

<On Organization Letter Head>

ANNEXURE-K

<EMD Declaration Form>

Tender No. :-

Date:-

To
Registrar
IIT Ropar

I/We/ M/s _____(Name) bid for the tender _____(Tender No.) towards
Supply/Service/of _____(Name of the items/service) declare that if we withdraw or modify our bids either
during the period of validity of bid or fail to execute the contract on award the purchase order/contract, we understand and agree that our firm
will be debarred for the period of one- year for further bidding of any tender of your Institute. Further, we agree that your Institute is at liberty
to intimate this debarment to all Departments/organizations of government and governmental organizations.

Signature of Bidder/ Agent

Name: _____

Designation: _____

Organization Name: _____

Contact No. : _____

<TO BE PROVIDED BY OEM ON LETTERHEAD>

DECLARATION OF COUNTRY OF ORIGIN

(To be given on Company Letter Head – For tender value below Rs.10 Crores)

(To be given by Statutory Auditor/ Cost Auditor/ Cost Accountant/ CA for tender value above Rs.10 Crores)

To,
The Registrar,
Indian Institute of Technology Ropar,
Rupnagar, Punjab - 140001

Subject: - Declaration of Country of Origin

Tender Reference No:

Name of Tender/ Work:

1. Country of Origin of Goods being offered: _____ (OM No. 6/18/2019-PPD dated 23.07.2020)

“Local Content” means the amount of value added in India which shall, be the total value of the item being offered minus the value of the imported content in the item (including all customs duties) as a proportion of the total value, in percent.

“*False declaration will be in breach of Code of Integrity under Rule 175(1)(i)(h) of the General Financial Rules for which a bidder or its successors can be debarred for up to two years as per Rule 151(iii) of the General Financial Rules along with such other actions as may be permissible under law.”

Yours faithfully,

(Signature of the bidder, with Official Seal)

Bid Submission

Online Bid Submission:

Envelope – 1
(Following documents to be provided as single PDF file)

| Sl. No. | Document | Content | File Types |
|----------------|-----------------|---|-------------------|
| 1 | Technical | Technical supporting documents in support of Technical Specifications Annexure A | |
| 2 | | Technical Compliance Sheet as per Annexure B | .pdf |
| 3 | | Manufacturer's Authorization Certificate as per Annexure C | pdf |
| 4 | | Non-blacklisting Certificate as per Annexure D | .pdf |
| 5 | | User List as per Annexure E | .pdf |
| 6 | | Warranty Certificate as per Annexure F | .pdf |
| 7 | | Bank Details as per Annexure G | .pdf |
| 8 | | Declaration as per Annexure H | .pdf |
| 9 | | Certificate as per Annexure I | .pdf |
| 10 | | Certificate as per Annexure J | .pdf |
| 11 | | EMD Declaration Form as per Annexure K | .pdf |
| 12 | | Certificate as per Annexure L | .pdf |
| 13 | | Any other documents in support of the abovesaid | .pdf |
| Sl. No. | Document | Content | File Types |
| 1 | Financial | Price bid should be submitted in given pricebid.xls format. (Note: -Comparison of prices will be done ONLY on the bids submitted for the Main Equipment and anything asked as 'Optional' in the specs is not to be included for overall comparison.) Bids for optional items are to be submitted in 'sheet2 Quote for optional items' | .xls |